

For Immediate Release: 3 September 2015


**INTERTEK WARNS OF COUNTERFEIT IEC/IECEE TEST CERTIFICATE (CB SCHEME)
IDENTIFYING BRACCO DIAGNOSTIC CARDIOGEN-82 INFUSION SYSTEM**

Intertek Testing Services notifies consumers, retailers and distributors that CB Test Certificate US/3702/ITS noting Bracco Diagnostics CardioGen-82 Infusion System associated to Report number 100216703ATL-001 dated June 18, 2013 is invalid and has been determined to be counterfeit.

These products have not been evaluated by Intertek. It is unknown if these products are compliant with the indicated international safety standard.

Counterfeit Document: IECEE CB Test Certificate US/3702/ITS
Applicant: Bracco Diagnostics
Product: Automatic Rubidium-82 Fluid Infusion and Irradiation System, Model 501

This notice only applies to the CB Test Certificate below identifying product Cardiogen-82 Infusion Systems model 501.

	IEC SYSTEM FOR CONFORMITY TESTING TO STANDARDS FOR SAFETY OF ELECTRICAL EQUIPMENT (IECEE) CB SCHEME		Ref. Certificate No. US/3702/ITS
	CB TEST CERTIFICATE		
Issued By:	Intertek Testing Services, N.A.		
Registration date:	2013-06-18		
Product:	Automatic rubidium-82 fluid infusion and irradiation system		
Applicant:	Bracco Diagnostics	250 Prospect Plains Road Monroe Township, NJ 08831	United States Of America
Manufacturer:			United States Of America
Factory:			United States Of America
Rating and principal Characteristics:	100/120/220/230-240 V, 8.3 A, 50/60 Hz. The 240 V setting on the appliance inlet is also used for 230 V.		
Trade mark (if any):			
Model/Type reference:	510		
Additional information:			
Sample of product tested to Be in conformity with IEC:	60601-1(ed.3);am1;am2		
Test Report Ref. No:	10021073-ATL-001		
INVALID			
This CB Test Certificate is issued by the National Certification Body Intertek Testing Services, N. A. 165 Main Street, Cortland, NY 13045, USA			
Signed by William T. Starr Date of issue: 2013-06-18		